



# Type Approval Certificate

This is to certify that the undernoted product(s) has/have been tested in accordance with the relevant requirements of the GL Type Approval System.

Certificate No. 82 784 - 92 HH

Company Tectis GmbH  
Carl-Legien-Straße 40-44  
63073 Offenbach, GERMANY

Product Description Pressure Gauges with Bourdon Tube

Type P 2324, P 2328  
P 2325, P 2329

Environmental Category None

Technical Data / Range of Application TECHNICAL DATA

Type:	P 2324 / P 2328	P 2325 / P 2329
Liquid filled:	None	Glycerine
Type of protection:	IP 54	IP 65
Nominal size:	100	100
Accuracy:	Class 1.0	
Indicating range:	- 0,6 ..... 0 bar up to 0 ..... 1000 bar	
Application range:	up to scale end value Dynamic 0,9 of scale end value	
Temperature range:	Ambient: - 25 °C to + 60 °C Medium: + 80 °C max	
Casing:	1.4301, Dial: Aluminium, Meter rectifier: 1.4571	
Pointer:	Aluminium, Segment gear: 1.4301/1.4305	

### RANGE OF APPLICATION

The pressure gauges of type P2324, P2325, P2328 and P2329 are type approved for application in piping systems for gaseous and liquid media according to the instructions of the manufacturer.

Test Standard Regulations for the Performance of Type Tests on Mechanical Components and Equipment, Part D.18 of Germanischer Lloyd. EN 837-1.

Documents - Data sheet DD 232 i, Drawing 2 324, 2 328, 2 325, 2 329  
- Test report No. U 348/92 dated 25.03.1992

Remarks None

Valid until 2018-04-28  
Page 1 of 1  
File No. II.B.01  
Hamburg, 2013-06-27

Type Approval Symbol



Germanischer Lloyd

Michael Kämpf

Hagen Markus